

Accuracy and reproducibility of trace and selected major element measurements in geological standard rocks using inductively coupled plasma mass spectrometry (ICP-MS)

Takashi Sano^{1*}, Yukiyasu Tsutsumi¹ and Arran P. Murch¹

¹ Department of Geology and Paleontology, National Museum of Nature and Science, 4-1-1 Amakubo, Tsukuba, Ibaraki 305-0005, Japan

*Author for correspondence: sano@kahaku.go.jp

Abstract We report the accuracy and reproducibility of 37 trace and 9 selected major element concentrations in 12 geological standard rocks using an inductively coupled plasma mass spectrometer. The geological standard rocks of the Geological Survey of Japan reference materials (JB-1, JB-1a, JB-1b, JB-2, JB-3, JA-1, JA-2, JA-3, JR-2, JSd-1) and the United States Geological Survey reference materials (BIR-1a, BHVO-2), were digested using HF-HClO₄-HNO₃, diluted to appropriate levels, and analyzed. The reproducibility of replicate analyses (i.e., measurement of several different solutions of the same reference material) is better than 5% (1 σ) for almost all elements. However, for Li, Sc, Ni and Zn, reproducibility is comparatively worse (>8%), likely caused by inhomogeneous distribution of the elements in the rock powder and/or interferences due to polyatomic ions. Our results are in agreement with reference values within 7% difference, except for Cr (up to 8% difference), Y and Zr (up to 10% difference) in some reference materials.

Key words: ICP-MS, volcanic rock, sediment, whole rock analysis

Introduction

Today, quantitative determinations of major and trace elements in rock samples are mostly performed by X-ray fluorescence spectrometry (XRF) and inductively coupled plasma mass spectrometry (ICP-MS). XRF measurements of rock samples provide relatively poor detection limits for some of the trace elements of interest and require correction of spectral overlaps for important elements such as Y, Zr and Nb (e.g., Sano *et al.*, 2020). Compared to XRF, the advantages of ICP-MS are its multi-element capability (e.g., applicable to all the naturally occurring rare-earth elements, REEs), low detection limits, and the simple spectra of the elements. Developments in analytical techniques have been made during the last two decades to obtain accurate and precise trace element abundances at ng/g (ppb) levels in rock samples using ICP-MS (e.g., Chang *et al.*, 2003; Nakamura and Chang, 2007; Senda *et al.*, 2014).

It is evident that when a method of analysis gains popularity, to guarantee the quality of the data pro-

duced, it is necessary to report the accuracy and reproducibility of analytical data in individual laboratories. We have published several papers to report trace element concentrations in rock digest solutions since 2016 (Sano *et al.*, 2016), however, major elements were not targeted. This paper, therefore, reports the establishment of a quantitative analytical routine for 37 trace and 9 major elements in volcanic rocks and sediment samples using ICP-MS. We also report the accuracy, reproducibility, and detection limits for each element based on several repeat analyses of 12 international reference materials and blank solutions during the last 7 years since 2014.

Analytical methods

Reagents, labware and calibration

The purified water (18.2 MQ cm) used throughout the present analysis was produced using a Milli-Q water purification system (Merck Millipore, Japan). Reagents used for sample digestion and dilution of standard solutions were TAMAPURE-AA grade 38% HF (AA-100), 70% HClO₄ (AA-100), and 68% HNO₃ (AA-100) (Tama Chemical Co, Ltd., Japan). 23-ml Teflon PFA vials with screw

caps (Savillex, USA) were used for sample decomposition. Two-types of polyethylene bottles, 10 ml (Yamayu Co Ltd., Japan) and 15 ml (Thermo Scientific Nalgene, USA), were used to store the sample solutions after decomposition and dilution, prior to ICP-MS analysis. To avoid contamination and to achieve low blanks, new polyethylene bottles were filled with 0.5 M HNO_3 (EL 1.38, Kanto Chemical Co. Inc., Japan) for more than 2 weeks and washed twice with Milli-Q water prior to use. PFA vials were rinsed prior to each use by fluxing with 6 M HCl (EL, Kanto Chemical Co. Inc., Japan) on a hot plate at 100°C for 3 days, followed by cleaning on a hotplate for another day with 5–10% TMS (Tama Chemicals Co, Ltd., Japan), followed by cleaning on a hotplate for 2 days with the 6 M HCl, followed by cleaning on a hotplate for 2 days with 6 M HNO_3 (EL 1.38, Kanto Chemical Co. Inc., Japan), and finally rinsed in Milli-Q water on a hotplate for another day. The vials were then filled with the 0.5 M HNO_3 and stored before the use.

The standard solutions for making calibration lines were prepared by gravimetric serial dilution from commercial standard solutions: a multi-element standard solution containing 10 µg/g (ppm) of REEs, Th and U (SPEX CertriPrep Int., USA); single element standard stock solutions containing 1,000 µg/g (ppm) for trace elements (Li, Be, V, Cr, Co, Ni, Cu, Zn, Ga, Rb, Sr, Y, Zr, Nb, Cs, Ba, Ce, Hf, Ta, Tl, Pb; SPEX CertriPrep Int., USA); and single element standard stock solutions containing 1,000 µg/g (ppm) for major elements (Ti, Al, Fe, Mn, Mg, Ca, Na, K and P; FUJIFILM Wako Pure Chemical Industries Ltd., Japan). The standard solutions were diluted by 2% HNO_3 for all elements excluding high-field strength elements (HFSE: Zr, Nb, Hf, Ta), and by 2% HNO_3 –0.1% HF for the HFSE. In order to reduce memory effects, multi-element standard solutions were prepared such that elements were present at levels no higher than those expected to occur in real samples after dilution. We prepared four or five diluted standard solutions including a reagent blank solution for each element (Table 1).

Sample digestion

Before sample digestion, 0.5 M HNO_3 was removed from the PFA vials and they were cleaned




Fig. 1. Analytical protocol for the acid digestion method.

twice with Milli-Q water. Then, the vials were dried on a hot plate at 80°C for ~1 hour. Samples were digested using the method shown in the analytical flow chart in Fig. 1. We digested 16 samples in one cycle.

Powdered sample (50–55 mg) was weighed into the vial, followed by 666 µl HClO_4 and 2,000 µl HF. We note that HClO_4 was added prior to HF to avoid violent reactions and the formation of fluorides in acid digestion. The vial was tightly capped and the powdered sample dissolved using an ultrasonic cleaning machine for 20 minutes. The sample vial

Table 1. Isotopes, cell gas mode, potential interference, calibration solutions, and count time to determine each element concentration. Detection limit and blank for each element are also shown.

Isotope	Cell gas mode	Potential Interferences	Calibration solutions (ng/g)	Count time (sec)	*DL (pg/g)	**Blank (pg)
⁷ Li	no gas		0, 0.001, 0.01, 0.1, 1	0.3	85	20
⁹ Be	no gas		0, 0.001, 0.01, 0.1, 1	3.0	1.1	0.93
²³ Na	no gas		0, 3, 30, 300, 3000	0.1	##n.d.	n.d.
²⁴ Mg	no gas	²³ Na ¹ H	0, 5, 50, 500, 5000	0.1	n.d.	n.d.
²⁷ Al	no gas	¹⁴ N spread	0, 10, 100, 1000, 10000	0.1	n.d.	n.d.
³¹ P	no gas	¹⁴ N ¹⁶ O ¹ H	0, 0.5, 5, 50, 500	0.3	n.d.	n.d.
³⁹ K	no gas	³⁸ Ar ¹ H	0, 1, 10, 100, 1000	0.1	n.d.	n.d.
⁴⁴ Ca	no gas	¹⁴ N ₂ ¹⁶ O	0, 10, 100, 1000, 10000	0.3	n.d.	n.d.
⁴⁵ Sc	no gas	⁴⁴ Ca ¹ H	0, 0.001, 0.01, 0.1, 1	0.1	67	110
⁴⁵ Sc	He	⁴⁴ Ca ¹ H	0, 0.001, 0.01, 0.1, 1		49	54
⁴⁷ Ti	no gas	³¹ P ¹⁶ O	0, 1, 10, 100, 1000	0.1	n.d.	n.d.
⁵¹ V	no gas		0, 0.2, 2, 20	0.3	340	520
⁵² Cr	no gas	³⁶ Ar ¹⁶ O	0, 0.2, 2, 20	0.1	26	44
⁵³ Cr	no gas	³⁶ Ar ¹⁶ O ¹ H	0, 0.2, 2, 20	0.1	610	330
⁵⁵ Mn	no gas	⁴⁰ Ar ¹⁴ N ¹ H, ³⁹ K ¹⁶ O	0, 0.5, 5, 50, 500	0.1	n.d.	n.d.
⁵⁶ Fe	no gas	⁴⁰ Ar ¹⁶ O, ⁴⁰ Ca ¹⁶ O	0, 10, 100, 1000, 10000	0.1	n.d.	n.d.
⁵⁹ Co	no gas	⁴³ Ca ¹⁶ O, ⁴⁰ Ar ¹⁸ O ¹ H	0, 0.3, 3, 30	0.3	140	210
⁶⁰ Ni	no gas	⁴⁴ Ca ¹⁶ O	0, 0.1, 1, 10	0.3	680	220
⁶³ Cu	no gas	⁴⁹ Ti ¹⁶ O, ²⁵ Mg ⁴⁰ Ar	0, 0.1, 1, 10	0.1	170	240
⁶⁶ Zn	no gas	²⁶ Mg ⁴⁰ Ar, ⁵⁰ Cr ¹⁶ O, ⁵⁰ Ti ¹⁶ O	0, 0.1, 1, 10	0.3	6200	3500
⁶⁶ Zn	He	²⁶ Mg ⁴⁰ Ar, ⁵⁰ Cr ¹⁶ O, ⁵⁰ Ti ¹⁶ O	0, 0.1, 1, 10		2272	440
⁷¹ Ga	no gas		0, 0.001, 0.01, 0.1, 1	0.3	20	31
⁸⁵ Rb	no gas		0, 0.1, 1, 10	0.1	36	48
⁸⁸ Sr	no gas		0, 0.2, 2, 20	0.1	87	49
⁸⁹ Y	no gas		0, 0.004, 0.04, 0.4, 4	0.3	9.3	12
⁹⁰ Zr	no gas	⁵⁰ Ti ⁴⁰ Ar, ⁵⁰ Cr ⁴⁰ Ar, ⁵¹ V ⁴⁰ Ar	0, 0.1, 1, 10	0.1	300	530
⁹³ Nb	no gas		0, 0.01, 0.1, 1	0.9	26	4.3
¹¹⁵ In#	no gas			0.1	n.d.	n.d.
¹¹⁵ In#	He				n.d.	n.d.
¹³³ Cs	no gas		0, 0.001, 0.01, 0.1, 1	1.5	0.61	0.56
¹³⁷ Ba	no gas		0, 0.2, 2, 20	0.3	43	50
¹³⁹ La	no gas		0, 0.001, 0.01, 0.1, 1	0.9	1.2	0.99
¹⁴⁰ Ce	no gas		0, 0.003, 0.03, 0.3, 3	0.9	3.1	3.4
¹⁴¹ Pr	no gas		0, 0.001, 0.01, 0.1, 1	2.1	0.80	0.48
¹⁴⁶ Nd	no gas		0, 0.001, 0.01, 0.1, 1	2.1	1.3	1.5
¹⁴⁷ Sm	no gas		0, 0.001, 0.01, 0.1, 1	2.1	0.89	0.61
¹⁵³ Eu	no gas	¹³⁷ Ba ¹⁶ O	0, 0.001, 0.01, 0.1, 1	2.1	0.64	0.26
¹⁵⁷ Gd	no gas	¹⁴⁰ Ce ¹⁶ O ¹ H, ¹⁴¹ Pr ¹⁶ O	0, 0.001, 0.01, 0.1, 1	2.1	1.4	1.4
¹⁵⁹ Tb	no gas		0, 0.001, 0.01, 0.1, 1	2.1	0.79	0.34
¹⁶³ Dy	no gas		0, 0.001, 0.01, 0.1, 1	2.1	0.64	0.47
¹⁶⁵ Ho	no gas		0, 0.001, 0.01, 0.1, 1	2.1	0.78	0.38
¹⁶⁶ Er	no gas		0, 0.001, 0.01, 0.1, 1	2.1	0.60	0.44
¹⁶⁹ Tm	no gas		0, 0.001, 0.01, 0.1, 1	2.1	0.66	0.38
¹⁷² Yb	no gas		0, 0.001, 0.01, 0.1, 1	2.1	0.72	0.51
¹⁷⁵ Lu	no gas		0, 0.001, 0.01, 0.1, 1	2.1	0.88	0.35
¹⁷⁸ Hf	no gas		0, 0.01, 0.1, 1	2.1	7.4	12
¹⁸¹ Ta	no gas	¹⁸⁵ Ho ¹⁶ O	0, 0.01, 0.1, 1	3.0	6.1	6.7
²⁰⁵ Tl	no gas		0, 0.001, 0.01, 0.1, 1	2.1	2.3	3.3
²⁰⁸ Pb	no gas		0, 0.001, 0.01, 0.1, 1	2.1	39	41
²⁰⁹ Bi#	no gas			0.3	n.d.	n.d.
²⁰⁹ Bi#	He				n.d.	n.d.
²³² Th	no gas		0, 0.001, 0.01, 0.1, 1	2.1	3.3	2.9
²³⁸ U	no gas		0, 0.001, 0.01, 0.1, 1	2.1	2.4	1.9

* DL (Detection limit): Three-sigma standard deviation of measurements of blank solutions. DL > 100 pg/g is shown in red.

** Blank: maximum values among 33 blank solutions. Blank > 100 pg is shown in red.

In and Bi are analyzed for internal standards.

n.d.: not determined for major and internal standard elements.

was then placed in an oven for 15 hours at 150°C and additional 20 hours at 160°C to fully dissolve the sample. The vial cap was then opened to dry the

sample solution by step heating at progressively increasing temperatures up to 160°C (Fig. 1). After this, 1,000 µl of HClO₄ was added and the vial was

Table 2. Instrument operating parameters.

Instrument	Agilent 7700x
Plasma power	1550 W
Torch	Quartz glass torch with Pt injection
Nebliser	Micro Flow Nebliser with I-AS, made of PFA
Spray chamber	Scott (double-pass) type, made of PFA
Cones	Nickel-plated copper
Plasma Ar gas flow rate	15 L/min
Carrier gas flow rate	0.8 L/min
Make-up Ar gas flow rate	0.3-0.6 L/min
Sample uptake rate	0.1 mL/min
Sampling depth	8.0 mm
Reaction mode	no gas (all elements) / + He gas (Sc, Zn)
He gas flow rate	4.0 mL/min (reaction mode)
Scan mode	peak jump
Acquisition time	71 sec × 3
Wash time	60 sec 2% HNO ₃ and 10 sec milli-Q water
Internal standard	¹¹⁵ In, ²⁰⁹ Bi
Oxide ratio (156 : 140)	< 1.5%

tightly capped and placed in an oven at 160°C for 12 hours. Next, the sample solution was refluxed with 2,000 μ l of 6 M HNO₃ and heated at 80°C for 1 hour. At this stage, no visible undissolved residue or floccule remained in the solution. Once again, the cap was opened to dry the sample solution by step heating at progressively increasing temperatures up to 160°C (Fig. 1). Finally, the dried sample was refluxed with 2% HNO₃–0.1% HF in the 15 ml polyethylene bottle to make a ‘mother sample solution’. Dilution factors at this stage were about 300.

Further dilution was undertaken with 2% HNO₃–0.1% HF in a 10 ml polyethylene bottle and ~100 ng/g (ppb) ¹¹⁵In and ²⁰⁹Bi solution was added as an internal standard before ICP-MS analysis. The internal standard was utilized to correct both instrumental drift and sensitivity drift during analysis of each sample solution. The drifts of low mass elements (Li to Nb) were normalized by In, those of high mass elements (Th and U) were normalized by Bi, and those of middle mass (Cs to Pb) were normalized by both In and Bi. Since volcanic rocks and sediments generally have no detectable In, this is the best element to use as an internal standard (Pretorius *et al.*, 2006). However, volcanic rocks and sediments sometimes have detectable Bi (e.g., JR-2). In this case, the drifts were normalized by only In. Total dilution factors were calculated for samples: 20,000–30,000 for JB-1, JB-1a, JB-1b, JB-3 and BHVO-2; 10,000–15,000 for JB-2, JA-1, JA-2, JA-3, JR-2 and JSd-1; 5,000–7,000 for BIR-1a.

A blank solution was also prepared in each cycle of sample digestion. The blank solution was ana-

lyzed after the sample solutions in order to monitor background memory effects during the course of the cycle. The above sample digestion procedure is itemized in the Appendix in detail.

Instrument

The ICP-MS used in this study is a quadrupole Agilent 7700x instrument (Agilent Technologies Japan, Ltd., Japan) at the National Museum of Nature and Science (NMNS). A HF-resistant spray chamber and PFA microflow nebulizer were employed to enable direct introduction of samples in diluted HNO₃–HF solution to the plasma since HFSE are unstable in simple HNO₃ solutions. All the solution introduction parts were carefully cleaned before use. Nickel-plated copper sample and skimmer cones were used throughout the study. The instrument parameters and operating conditions are shown in Table 2.

The instrument parameter was optimized daily using a 1 ng/g (ppb) tuning solution (Agilent Technologies Japan, Ltd., Japan) containing the whole mass range of the elements from Li to Tl (⁷Li, ²⁴Mg, ⁵⁹Co, ⁸⁹Y, ¹⁴⁰Ce, ²⁰⁵Tl) to obtain maximum signal intensity, minimum oxide formation rates and required counts (>10,000, >150,000, >15,000, >20,000, >20,000 and >10,000 counts/sec for Li, Mg, Co, Y, Ce and Tl, respectively). Signals of all elements were counted by pulse mode (for low count rates) and analog mode (for high count rates). Sensitivity differences of the two modes were optimized at least once per month using a 10 ng/g (ppb) tuning solution (Agilent Technologies Japan, Ltd.,

Japan) containing the whole mass range of elements. Oxide formation level was maintained at less than 1.5% of CeO relative to Ce, but we had to correct for overlaps between the spectra of oxides and hydroxides for some REEs (Table 1). To overcome this problem, we measured three separate solutions prepared by 1 ng/g (ppb) REEs solution, 5 ng/g (ppb) Ce, and 10 ng/g (ppb) Ba with 2 ng/g (ppb) Pt (SPEX CertriPrep Int., USA), and correction factors for the oxides and hydroxides were calculated (<1.5%) as follows.

$$^{157}\text{Gd}: ^{157}\text{Gd} - 0.00145 \times ^{140}\text{Ce} - 0.0106 \times ^{141}\text{Pr}$$

$$^{153}\text{Eu}: ^{153}\text{Eu} - 0.00156 \times ^{137}\text{Ba}$$

$$^{181}\text{Ta}: ^{181}\text{Ta} - 0.0033 \times ^{165}\text{Ho}$$

These correction factors were applied to all measurements. The interference due to polyatomic ions affect measurements of trace elements as listed in Table 1 (May and Wiedmeyer, 1998; Pretorius *et al.*, 2006), but we did not conduct any correction for other trace elements in this study. Only Argon was used as makeup gas and carrier gas for the measurements of all elements (no gas mode), but we also tried to introduce Helium collision mode (He mode) for measurements of Sc and Zn that have high background to signal intensity (Table 1).

Results and Discussion

Detection limit and procedural blank

Detection limits (DLs) for trace elements were calculated using the results of analyses of 33 blank solutions that were separately digested over the last 7 years. The DLs are defined as three times the standard deviation of the analytical results. Table 1 shows that the DLs of most REEs are lower than 1 pg/g (ppt). These DLs theoretically allow the determination of sub ng/g (ppb) level elements in rock samples, even using a highly distributed sample solution in this study. Consequently, the procedural blank for a trace element is the main limitation in such highly sensitive ICP-MS analysis. In contrast to the REEs, some other trace elements (Cr, Co, Ni, Cu, Zn, Zr) have relatively high DLs (>100 pg/g: ppt), which were certainly caused by interference effects (Table 1). The slightly higher DLs of Rb, Sr, Ba and Pb (>30 pg/g: ppt) are probably due to contamination of these elements during preparation of blank solutions. The maximum blank

for the trace elements is nanogram level (3.5 ng for Zn). We conclude that precise quantifications are possible for the analyzed elements in Table 1 because their concentrations in general volcanic rocks and sediments are usually at $\mu\text{g/g}$ (ppm) level (Table 3). The high DLs for some elements suggest that the precise determinations of sub $\mu\text{g/g}$ (ppm) level are sometimes difficult for these elements. For example, the procedural blank of Co (210 pg) is about 3% of the Co content (0.14 $\mu\text{g/g}$: ppm) in a 50 mg sample powder for JR-2 (Table 3). We should therefore consider the DLs and procedural blanks when we analyze elements with sub $\mu\text{g/g}$ (ppm) levels (e.g., REEs in ultramafic rocks).

Reproducibility and Comparison to reference values

Over the last 7 years, we performed 26 separate digestions and trace element analyses for JB-1a (Table 3), and 16 solutions among them were selected for major element analysis by ICP-MS (Table 4). Multiple digestions and analyses were also conducted for another 11 geological standard rocks (Tables 3 and 4). The listed relative standard deviation (RSD) for each element in Tables 3 and 4 are likely to adequately account for uncertainties arising from weighing, digestion, and sample and calibration solution dilution and preparation, accounting for the fact that the replicated digestions were randomly performed over a period of 7 years by several scientists.

The reproducibility of replicated analyses (i.e., measurement of several different solutions of the same reference material) in the 12 geological standard rocks is better than 5% RSD for almost all elements (Table 3). However, the reproducibility for Li, Sc, Ni and Zn in many samples, is worse than 8% (red in Table 3). Interferences due to polyatomic ions for Sc, Ni and Zn (Table 1) are likely the cause of the relatively high RSD. The relatively high RSD for Li suggests this element may be affected by inhomogeneous distribution in the rock powders. We also note that the RSDs in sub $\mu\text{g/g}$ (ppm) concentrations (e.g., Ta, Tl, Th and U in BIR-1a) are generally high caused by their high procedural blanks compared to their concentrations.

Measured concentrations of most major and trace elements in the 12 geological standard rocks agree with the reference (preferred, recommended, certi-

Table 3. Trace element concentrations ($\mu\text{g/g}$: ppm) in reference rocks.

Sample	Run No.	Li	Be	Sc	V	Cr	Co	Ni	Cu	Zn	Ga	Rb	Sr	Y	Zr	Nb	Cs	Ba	La	Ce
JB-1	EG11	9.4	1.20	26.0	206	\$	35.0	128	51		17.0	37.3	437	19.5	126	33.1	1.33	482	35.39	63.37
	AF11	10.7	1.38	28.0	220		37.5	136	56	82	17.8	37.4	441	20.0	126	36.5	1.37	492	37.22	66.30
	AF23	10.5	1.28	26.3	216		36.3	135	53	79	17.2	36.5	429	19.5	124	31.9	1.31	470	35.56	63.18
	KH34	10.3	1.30	26.0	193	461	35.8	132	54	88	17.4	37.3	430	20.1	130	31.0	1.29	487	35.43	63.44
	KH58	10.5	1.38	27.3	225		38.7	143	57	77	18.6	38.9	449	21.6	137	37.6	1.38	507	37.82	67.60
	KH70	10.1	1.27	25.4	194	451	35.4	131	53	60	17.1	37.5	427	20.2	129	34.5	1.31	473	35.65	63.98
	OR22	10.4	1.27	26.0	196	448	34.7	124	53	77	17.2	33.1	405	18.0	111	31.7	1.28	455	34.72	61.90
	KM14-2	12.0	1.48	27.2	200	462	36.8	126	48	68	17.7	37.8	426	20.8	128	33.9	1.36	482	36.58	65.10
	ES16	10.2	1.27	27.8	221	464	37.2	139	54	98	17.9	38.7	446	22.0	134	36.7	1.34	498	42.37	74.32
	ET48	8.6	1.13	27.5	221	465	35.4	130	49	87	17.3	38.0	430	22.3	141	36.6	1.35	498	45.59	78.23
JB-1	average	10.3	1.30	26.8	209	459	36.3	132	53	80	17.5	37.2	432	20.4	129	34.4	1.33	484	37.63	66.74
	RSD*	0.9	0.10	0.9	13	7	1.2	6	3	11	0.5	1.6	12	1.3	8	2.4	0.04	15	3.55	5.36
	RSD (%)**	8.3	7.6	3.4	6.1	1.6	3.4	4.6	5.4	14.3	2.8	4.3	2.9	6.4	6.4	7.0	2.6	3.2	9.4	8.0
JB-1	pref.#	11.5	1.33	29	211	430	38.1	140	56.7	85.2	17.9	39	444	24	142	28	1.2	489	38	66
JB-1a	EG23	10.3	1.32	28.3	203		37.0	129	56	80	17.4	38.4	443	21.6	137	26.0	1.23	495	36.99	65.89
	KF36	11.8	1.52	34.1	241	407	39.6	150	60	69	18.7	42.8	465	21.5	158	26.5	1.16	460	34.56	61.45
	PP11	11.6	1.51	30.2	234	409	38.8	146	57	83	18.5	43.0	462	21.4	134	26.3	1.16	459	34.47	61.05
	OJ47	7.7	1.33	28.2	208	453	36.1	151	54	77	17.4	37.4	438	21.2	135	25.5	1.18	482	35.60	62.92
	OJ75	10.6	1.31	27.9	202		37.0	141	55	87	17.7	37.4	459	20.9	137	25.6	1.19	486	35.05	62.72
	E15	13.1	1.47	31.6	222		39.7	144	59	86	18.7	39.4	421	21.9	136	25.8	1.21	478	35.29	62.77
	OJ90	11.1	1.34	28.1	220	424	37.4	139	54	67	17.7	37.3	446	20.8	135	25.6	1.18	476	34.92	62.24
	PP27	11.7	1.38	26.8	194	417	36.1	131	49	71	17.4	35.9	435	20.5	130	23.9	1.15	473	34.84	62.18
	M23	10.7	1.41	28.4	198		36.4	133	46		17.3	37.3	450	21.5	137	25.8	1.26	500	36.65	65.19
	KI14	11.6	1.44	29.2	232	432	39.0	146	56	88	18.7	39.1	458	21.4	142	26.9	1.22	495	36.81	64.62
	MG07	11.4	1.40	29.1	235		39.0	148	57	111	18.5	38.6	460	21.4	139	26.5	1.17	475	35.31	63.00
	OJ90m	11.2	1.36	27.5	192		37.2	135	53	89	17.5	36.5	449	21.1	135	25.4	1.21	489	36.58	64.82
	PA14	11.3	1.43	31.0	239	421	40.3	148	58	85	19.2	39.7	456	22.1	144	27.2	1.23	490	36.49	63.88
	A14	10.5	1.29	27.2	211	412	36.3	133	53	79	17.5	36.4	435	20.5	133	25.0	1.20	488	36.67	64.14
	AM14	11.5	1.42	28.5	216	416	37.4	178	56	75	18.0	35.7	448	20.9	137	25.6	1.19	488	36.43	64.23
	SA6	11.1	1.39	29.2	210	414	38.4	150	57		18.4	38.6	468	21.2	142	26.7	1.18	472	35.38	62.96
	A30	11.5	1.41	30.4	217	416	39.3	165	59		18.9	40.0	480	21.8	143	27.0	1.22	477	36.25	64.19
	SA16	11.1	1.33	28.9	218	414	36.8	140	54		18.0	36.4	453	19.7	132	25.8	1.19	497	36.71	65.29
	SA32	11.7	1.46	27.5	202		37.9	137	53	86	17.9	35.1	442	20.8	133	25.5	1.19	479	35.69	63.85
	A46	11.8	1.38	27.3	218	430	37.6	137	51	86	17.9	35.6	444	19.6	123	24.9	1.19	486	35.79	63.62
	PI14	8.7	1.08	22.1	168	419	32.4	117	45	85	16.0	33.5	423	18.7	117	23.5	1.21	499	36.55	65.16
	P30	10.7	1.29	25.7	206		36.0	134	50	78	17.4	35.6	432	19.7	125	23.8	1.17	479	35.67	63.55
	SA48	11.1	1.42	27.3	205	425	38.0	141	54	79	18.2	36.7	443	21.0	131	25.1	1.19	482	36.13	63.97
	SA64	12.0	1.47	27.9	208	414	38.3	140	55	79	18.3	36.7	450	21.1	131	25.4	1.18	483	36.20	64.32
	KM14	11.5	1.39	26.8	206	421	37.0	136	53	80	17.7	35.6	437	20.7	132	24.8	1.18	472	34.90	62.15
	EG51	12.4	1.45	29.9	234	433	38.3	144	56	87	17.7	37.9	444	20.7	133	25.4	1.17	473	34.66	61.78
JB-1a	average	11.1	1.38	28.4	213	421	37.6	142	54	82	18.0	37.6	448	20.9	135	25.6	1.19	482	35.79	63.54
	RSD	1.1	0.09	2.2	17	11	1.6	12	4	9	0.7	2.2	14	0.8	8	0.9	0.02	11	0.79	1.27
	RSD (%)	9.5	6.4	7.7	7.9	2.6	4.3	8.3	6.8	10.9	3.7	5.9	3.0	3.7	5.7	3.7	2.1	2.2	2.2	2.0
JB-1a	pref.	10.8	1.44	27.81	200.3	408	38.53	139.5	54.5	88.5	18.13	38.15	443.4	22.91	140.1	27.57	1.216	495.1	37.74	65.93
JB-1b	ET14	10.3	1.21	30.6	239	457	39.9	153	53	76	18.1	35.0	443	20.8	129	25.7	0.84	492	38.10	65.85
	ET27	9.2	1.12	28.9	215	455	36.2	138	48	74	16.9	34.1	431	20.2	125	24.2	0.85	504	39.39	67.76
	IZ11	11.7	1.37	28.7	229	448	38.7	150	51	84	17.5	33.7	436	19.2	118	24.7	0.80	479	36.95	64.95
	EG52	11.9	1.31	30.8	241	466	39.5	152	55	85	17.4	35.0	444	20.0	125	24.7	0.82	487	36.77	64.31
	average	10.8	1.25	29.8	231	456	38.6	148	52	80	17.5	34.5	439	20.1	124	24.8	0.83	490	37.80	65.72
JB-1b	RSD	1.3	0.11	1.1	12	7	1.6	7	3	5	0.5	0.6	6	0.7	4	0.6	0.02	10	1.21	1.50
	RSD (%)	12.1	8.7	3.7	5.1	1.6	4.3	4.8	6.0	6.8	2.8	1.9	1.4	3.4	3.6	2.6	3.0	2.1	3.2	2.3
	recom.##	10.8	1.3		214	439	40.3	148	55.5	80		39.1	439				1.21			
JB-2	KH02B	8.4	0.26	49.3	597		36.5	14.9	219	98	16.1	6.2	181	21.6	46.4	0.48	0.75	215	2.17	6.30
	EG24	7.4	0.24	54.1	571		36.2	11.0	219	114	16.1	6.4	171	22.2	46.1	0.44	0.78	214	2.17	6.38
	EG36	8.6	0.25	61.3			37.5	21.0	222	123	16.4	6.5	180	22.4	48.0	0.46	0.78	215	2.15	6.27
	OJ35	6.2	0.22	61.3	554	26.3	33.4	12.0	185	102	15.0	5.9	163	20.4	47.8	0.42	0.71	190	1.93	5.70
	KF37	8.4	0.27	64.9			38.2	14.3	200	121	16.7	7.1	186	22.6	53.4	0.48	0.76	201	2.07	6.04
	PP12	8.2	0.26	64.0		29.9	37.9	14.0	196	122	16.5	7.0	184	22.3	47.0	0.49	0.75	200	2.05	5.97
	OJ48	6.5	0.22	55.8	574	30.4	34.8	12.3	192	112	15.8	6.1	174	21.8	45.6	0.48	0.76	210	2.09	6.15
	OJ59	6.0	0.21	55.8	578	28.8	35.0	1												

Table 3 (continued). Trace element concentrations ($\mu\text{g/g}$; ppm) in reference rocks.

Sample	Run No.	Li	Be	Sc	V	Cr	Co	Ni	Cu	Zn	Ga	Rb	Sr	Y	Zr	Nb	Cs	Ba	La	Ce	
JA-1	KH04B	11.3	0.51	26.8	106		11.5	5.6	45	80	17.6	11.1	265	27.4	84.5	1.29	0.60	305	5.02	13.25	
	PP28	10.8	0.46	27.2	99	11.7	10.3	1.8	37	85	16.4	10.4	250	25.6	77.3	1.09	0.59	285	4.55	12.30	
	PA15	10.8	0.49	32.7	129	11.6	11.9	2.0	45	108	19.2	12.1	277	29.1	89.4	1.35	0.66	308	5.01	13.31	
	A15	9.6	0.43	27.1	105	10.2	10.2	2.0	39	92	16.5	10.6	247	26.0	79.4	1.17	0.62	294	4.85	12.88	
	ET28	9.2	0.42	27.9	107	9.3	10.8	3.0	39	90	16.7	10.8	253	26.5	79.7	1.16	0.63	302	4.96	13.04	
	A47	11.0	0.47	28.0	110	9.9	10.7	2.0	40	98	17.1	10.3	261	24.7	77.3	1.22	0.62	293	4.75	12.79	
JA-1	average	10.4	0.46	28.3	109	10.5	10.9	2.7	41	92	17.2	10.9	259	26.6	81.3	1.21	0.62	298	4.86	12.93	
	RSD	0.8	0.03	2.2	10	1.1	0.7	1.5	3	10	1.1	0.7	11	1.6	4.8	0.09	0.03	9	0.18	0.37	
JA-1	pref.	10.43	0.53	27.9	106.2	7.5	11.51	2.2	42.5	88.3	16.7	11.02	259.3	28	83.7	1.333	0.627	304	4.88	13.15	
JA-2	MG08	31.0	2.27	19.4	136		28.7	137	31		17.3	73.5	257	16.3	118	8.8	4.88	304	15.27	32.20	
	ET29	26.2	1.94	17.6	119	430	26.1	121	28	64	16.3	69.9	238	15.6	120	8.3	5.03	316	15.89	32.88	
	ET30	26.4	1.95	17.6	120	423	26.3	120	29	63	16.4	70.6	243	15.6	112	8.3	5.12	318	16.14	33.33	
	IZ13	32.9	2.40	18.0	131	422	28.3	130	30	69	17.2	71.0	252	15.1	116	8.6	4.92	308	15.29	32.22	
JA-2	average	29.1	2.14	18.1	127	425	27.3	127	29	65	16.8	71.2	248	15.7	117	8.5	4.99	312	15.65	32.66	
	RSD	3.4	0.23	0.9	9	4	1.3	8	2	3	0.5	1.6	8	0.5	3	0.3	0.11	7	0.44	0.55	
JA-2	pref.	29.18	2.26	18.93	119.7	425	28.33	136	29	64.5	16.85	69.8	245.8	16.89	108.5	9.3	4.78	308.4	15.46	32.86	
JA-3	KH05B	15.0	0.79	20.7	176		22.3	37.7	49	66	17.9	37.8	302	19.5	122	3.08	2.17	333	9.64	22.54	
	ET15	13.8	0.70	23.1	199	67	22.4	34.5	48	77	18.4	37.9	293	19.8	130	3.17	2.19	325	9.48	22.01	
	ET31	12.1	0.61	19.7	168	69	19.4	30.5	42	65	16.1	34.3	273	18.0	111	2.82	2.11	316	9.35	21.54	
	AM15	13.9	0.69	20.6	178		20.0		45	57	16.4	33.7	289	18.0	112	2.90	2.05	310	9.00	21.05	
	A31	13.7	0.68	21.8	194	71	21.1	39.5	46	51	17.3	36.9	296	18.8	117	3.00	2.08	299	8.86	20.76	
JA-3	average	13.7	0.70	21.2	183	69	21.0	35.6	46	63	17.2	36.1	290	18.8	118	2.99	2.12	316	9.26	21.58	
	RSD	1.0	0.06	1.3	13	2	1.3	3.9	3	10	1.0	1.9	11	0.8	8	0.14	0.06	13	0.33	0.72	
JA-3	pref.	14.5	0.80	22.0	169	66.2	21.1	32.2	43.4	68	16.3	36.7	287.0	21.1	118	3.4	2.08	323.0	9.33	22.80	
JR-2	KH06B	86.2	3.81	5.5	4.1		0.18		5.4	23	18.4	316		9.1	46.6	94.3	18.2	26.4	31.1	15.34	38.36
	IZ14	92.7	4.08	5.8	1.4	5.8	0.12	0.8	1.0	32	18.0	303		7.9	42.9	83.6	17.3	24.8	27.2	14.22	35.71
	P31	77.3	3.43	5.0	1.3	6.5	0.11	0.8	0.9	23	17.0	283		7.7	41.3	79.7	16.4	24.6	26.9	14.17	35.76
	EG54	96.5	3.95	6.1	1.5	3.7	0.14	0.7	1.3	28	17.9	308		8.1	44.6	87.7	17.0	24.5	27.4	14.01	35.46
JR-2	average	88.2	3.82	5.6	2.1	5.3	0.14	0.8	2.2	27	17.8	303		8.2	43.8	86.3	17.2	25.1	28.2	14.44	36.32
	RSD	8.4	0.28	0.4	1.3	1.4	0.03	0.1	2.2	4	0.6	14		0.6	2.3	6.2	0.7	0.9	2.0	0.61	1.36
JR-2	recom.	79.2	3.75	5.6	3.0	3.1	0.46	2.0	1.4	28	17.9	303.0		8.1	51.5	96.3	18.7	25.0	39.5	16.30	38.80
BIR-1a	KH07B	3.36	0.109	41.5	362		54.8	183	128	67	16.1	0.18	113	15.1	15.7	0.50		7.02	0.63	1.97	
	KH33	3.00	0.097	38.2	316		49.1	162	114	60	14.3	0.20	103	13.6	13.5	0.44		5.90	0.56	1.74	
	KH59	2.79	0.088	45.9	326	425	54.6	169	120	78	15.7	0.16	107	14.7	14.4	0.47	0.0062	6.47	0.60	1.91	
	KH71	2.90	0.089	40.5	325	414	50.9	167	118	63	14.9	0.21	104	14.0	13.9	0.47	0.0029	6.42	0.58	1.84	
	KH83	2.82	0.081	38.8	317		48.3	161	113		14.1	0.40	105	13.3	13.1	0.45	0.0076	5.78	0.55	1.72	
	OR11	2.68	0.085	45.0	293		49.7	154	104	62	13.8	0.25	97	12.8	11.8	0.20	0.0043	5.68	0.54	1.73	
	OR23	2.79	0.082	44.8	292	427	51.0	166	112	67	14.0	0.30	98	12.1	11.5	0.34	0.0038	5.80	0.55	1.74	
	AF12	2.75	0.085	45.7	311		50.1	162	110	71	14.2	0.23	99	12.8	12.4	0.39	0.0054	5.87	0.55	1.74	
	AF24	2.82	0.084	46.0	325	409	52.5	168	110	70	14.6	0.27	104	13.5	13.2	0.47	0.0044	6.14	0.58	1.82	
	EG12	2.76	0.090	44.0	332		52.0	171	114	77	15.0	0.29	110	14.0	13.9	0.49	0.0072	6.44	0.59	1.85	
	KI15	3.14	0.094	51.5	357	400	54.8	177	119	86	15.8	0.23	111	15.0	15.2	0.51	0.0065	6.38	0.60	1.86	
	BIR-1a	average	2.89	0.089	43.8	323	415	51.6	167	115	70	14.8	0.25	104	13.7	13.5	0.43	0.0054	6.17	0.57	1.81
	RSD	0.20	0.008	3.8	22	11	2.3	8	6	8	0.8	0.07	5	0.9	1.3	0.09	0.0016	0.41	0.03	0.08	
	RSD (%)	7.0	9.1	8.7	6.8	2.7	4.5	4.7	5.6	11.4	5.4	27.3	5.0	6.9	9.8	21.3	29.7	6.6	4.8	4.6	
BIR-1	pref.	3.203	0.102	43.21	320.6	392.9	52.22	168.9	120.7	70.4	15.46	0.21	108.6	15.6	14.8	0.553	0.00646	6.75	0.627	1.92	
BHVO-2	SA7	4.6	1.01	34.3		304	47.0	134	139		22.5	9.68	425	24.7	179	17.0	0.090	124	14.47	36.36	
	SA33	4.4	1.06	30.3	330	292	43.3	116	124	109	20.3	8.48	382	22.9	160	16.2	0.094	120	13.97	34.66	
	IZ15	5.0	1.07	31.9	353	306	45.9	123	133	116	21.6	8.80	404	22.9	160	16.3	0.091	124	14.13	35.49	
	SA49	4.4	1.01	31.1	326	307	44.6	119	130	103	21.4	9.04	390	23.8	162	15.9	0.089	126	14.46	36.25	
	SA65	4.7	1.09	31.9	345		45.6	120	131	105	21.7	8.95	400	24.1	165	16.7	0.096	125	14.60	36.62	
BHVO-2	average	4.6	1.05	31.9	339	302	45.3	122	132	108	21.5	8.99	400	23.7	165	16.4	0.092	124	14.33	35.88	
	RSD	0.2	0.04	1.5	13	7	1.4	7	5	6	0.8	0.44	16	0.8	8	0.5	0.003	2	0.27	0.80	
	RSD (%)	4.9	3.7	4.7	3.7	2.2	3.1	5.9	4.1	5.4	3.7	4.9	4.0	3.4	4.8	2.8	3.3	1.9	1.9	2.2	
BHVO-2	pref.	4.5	1.076	31.83	318.2	287.2	44.89	119.8	129.3	103.9	21.37	9.261	394.1	25.91	171.2	18.1	0.0996	130.9	15.2	37.53	
JSd-1	E16	26.7	1.40	11.6	81		10.8	6.9	22	94	17.5	67.2	316	13.0	21.9	10.1	1.96	507	15.49	31.08	
	SA17	22.4	1.20	10.3	72		9.9	9.4	21	77	16.2	61.9	328	12.5	23.6	9.8	1.96	513	15.13	30.23	
	PI15	19.2	1.06	8.7	64		9.2	6.4	20	93	15.0	58.7	322	11.5	20.1	9.3	1.98	531	15.84	31.54	

Table 3 (continued). Trace element concentrations ($\mu\text{g/g}$; ppm) in reference rocks.

Sample	Run No.	Pr	Nd	Sm	Eu	Gd	Tb	Dy	Ho	Er	Tm	Yb	Lu	Hf	Ta	Tl	Pb	Th	U
JB-1	EG11	6.76	25.28	4.88	1.40	4.56	0.73	3.98	0.78	2.24	0.32	2.02	0.30	3.48	2.36	0.095	6.47	7.99	1.39
	AF11	7.04	26.09	5.04	1.42	4.66	0.74	4.11	0.81	2.31	0.31	2.04	0.30	3.47	2.45	0.084	6.81	9.11	1.51
	AF23	6.68	24.88	4.83	1.38	4.48	0.71	3.89	0.77	2.19	0.30	1.94	0.29	3.40	2.16	0.086	6.75	7.91	1.39
	KH34	6.67	25.10	4.77	1.39	4.45	0.72	3.95	0.78	2.22	0.30	1.97	0.30	3.51	2.07	0.089	6.53	8.42	1.52
	KH58	7.13	26.44	5.11	1.54	5.13	0.77	4.18	0.81	2.32	0.31	2.10	0.30	3.65	2.65	0.086	6.51	8.33	1.51
	KH70	6.75	24.95	4.88	1.48	4.94	0.71	3.96	0.77	2.21	0.30	1.97	0.28	3.44	2.44	0.083	6.45	8.18	1.50
	OR22	6.54	24.51	4.71	1.39	4.61	0.71	3.92	0.76	2.14	0.31	1.97	0.29	3.27	2.18	0.095	7.15	10.36	1.69
	KM14-2	6.86	25.31	4.77	1.44	4.74	0.74	4.01	0.77	2.23	0.31	1.99	0.28	3.61	2.53	0.086	6.51	8.96	1.59
	ES16	7.36	26.32	5.01	1.47	4.40	0.80	4.00	0.81	2.34	0.31	2.01	0.31	3.61	2.55	0.094	6.13	9.34	1.66
ET48	7.57	26.73	5.00	1.46	3.97	0.80	4.09	0.81	2.33	0.32	2.03	0.31	3.83	2.72	0.092	6.09	9.99	1.68	
JB-1	average	6.93	25.56	4.90	1.44	4.59	0.74	4.01	0.79	2.25	0.31	2.00	0.30	3.53	2.41	0.089	6.54	8.86	1.54
	RSD	0.33	0.77	0.13	0.05	0.31	0.04	0.09	0.02	0.07	0.01	0.05	0.01	0.16	0.22	0.005	0.31	0.85	0.11
	RSD (%)	4.8	3.0	2.7	3.6	6.8	4.9	2.3	2.6	3.0	2.0	2.2	3.4	4.4	9.0	5.2	4.8	9.6	7.2
JB-1	pref.	7.2	26.3	5.1	1.46	4.7	0.72	4.1	0.83	2.2	0.3	2.16	0.31	3.5	1.8	0.087	6.3	9.2	1.6
JB-1a	EG23	6.87	25.43	4.91	1.40	4.35	0.71	3.93	0.77	2.19	0.31	1.97	0.29	3.51	1.57	0.079	5.61	7.87	1.41
	KF36	6.51	24.11	4.68	1.31	4.17	0.66	3.71	0.73	2.06	0.28	1.85	0.27	3.54	1.41	0.071	5.73	8.03	1.48
	PP11	6.47	23.78	4.58	1.29	4.09	0.65	3.68	0.72	2.04	0.28	1.80	0.27	3.13	1.38	0.073	5.70	8.12	1.46
	OJ47	6.70	24.78	4.79	1.35	4.18	0.68	3.89	0.76	2.14	0.30	1.92	0.29	3.45	1.53	0.081	6.20	8.93	1.55
	OJ75	6.69	24.81	4.83	1.35	4.29	0.69	3.89	0.76	2.15	0.30	1.94	0.29	3.44	1.52	0.076	6.04	8.62	1.55
	E15	6.69	24.65	4.81	1.34	4.52	0.67	3.91	0.74	2.11	0.28	1.91	0.26	3.23	1.45	0.063	6.35	8.46	1.55
	OJ90	6.63	24.51	4.77	1.37	4.49	0.71	3.88	0.77	2.18	0.30	1.94	0.29	3.37	1.49	0.084	6.00	8.68	1.55
	PP27	6.69	24.72	4.85	1.41	4.63	0.71	3.99	0.78	2.21	0.31	2.00	0.30	3.54	1.55	0.085	5.71	8.39	1.51
	M23	6.89	25.52	4.99	1.44	4.35	0.73	4.03	0.82	2.34	0.29	1.97	0.34	3.79	1.70	0.106	4.60	8.97	1.46
	KI14	6.83	25.23	4.87	1.36	4.36	0.71	3.94	0.79	2.18	0.31	2.00	0.30	3.45	1.54	0.086	5.86	9.02	1.58
	MG07	6.60	24.40	4.67	1.33	4.31	0.68	3.80	0.75	2.14	0.29	1.86	0.27	3.26	1.44	0.092	5.84	8.07	1.50
	OJ90m	6.85	25.49	5.03	1.37	4.44	0.71	4.09	0.79	2.25	0.32	2.11	0.31	3.58	1.60	0.089	6.12	8.79	1.58
	PA14	6.77	24.92	4.79	1.36	4.38	0.69	3.79	0.76	2.11	0.29	1.92	0.29	3.33	1.49	0.082	5.78	8.63	1.52
	A14	6.84	25.31	4.92	1.43	4.62	0.74	4.04	0.81	2.24	0.32	2.07	0.32	3.70	1.64	0.103	5.76	9.18	1.62
	AM14	6.84	25.30	4.92	1.40	4.49	0.73	4.03	0.80	2.26	0.31	2.03	0.30	3.68	1.64	0.092	5.46	8.24	1.48
	SA6	6.61	24.48	4.67	1.30	4.15	0.66	3.76	0.73	2.10	0.28	1.89	0.26	3.09	1.38	0.060	6.12	9.02	1.62
	A30	6.77	25.03	4.75	1.31	4.09	0.66	3.85	0.75	2.11	0.28	1.89	0.27	3.12	1.41	0.067	6.11	9.09	1.60
	SA16	6.95	25.78	5.00	1.45	4.79	0.75	4.13	0.82	2.31	0.32	2.11	0.31	3.78	1.69	0.089	5.93	8.87	1.59
	SA32	6.73	24.99	4.81	1.37	4.49	0.72	3.97	0.78	2.22	0.31	2.00	0.30	3.48	1.57	0.088	5.55	8.12	1.48
	A46	6.75	25.24	4.87	1.39	4.60	0.72	4.02	0.78	2.24	0.31	2.01	0.29	3.47	1.59	0.090	5.82	8.37	1.54
	PI14	7.00	26.20	5.05	1.48	4.92	0.77	4.31	0.85	2.42	0.34	2.17	0.33	3.94	1.75	0.108	6.18	9.21	1.70
	P30	6.72	25.06	4.82	1.37	4.38	0.71	4.02	0.78	2.21	0.31	1.99	0.29	3.48	1.42	0.083	5.69	8.48	1.53
	SA48	6.79	25.35	4.86	1.39	4.43	0.70	4.00	0.77	2.23	0.31	2.00	0.28	3.47	1.53	0.079	5.76	8.48	1.51
	SA64	6.79	25.35	4.87	1.38	4.42	0.70	4.02	0.77	2.21	0.30	1.99	0.29	3.47	1.58	0.078	5.81	8.61	1.53
	KM14	6.61	24.52	4.82	1.38	4.38	0.70	3.91	0.77	2.19	0.29	1.97	0.29	3.42	1.55	0.085	5.98	8.83	1.59
	EG51	6.54	24.34	4.70	1.32	4.28	0.69	3.80	0.75	2.10	0.31	1.90	0.28	3.31	1.49	0.086	5.80	8.20	1.50
JB-1a	average	6.74	24.97	4.83	1.37	4.41	0.70	3.94	0.77	2.19	0.30	1.97	0.29	3.46	1.53	0.084	5.83	8.59	1.54
	RSD	0.13	0.54	0.12	0.05	0.20	0.03	0.14	0.03	0.09	0.01	0.09	0.02	0.21	0.10	0.012	0.33	0.39	0.06
	RSD (%)	2.0	2.2	2.4	3.4	4.5	4.2	3.5	3.9	4.0	4.8	4.3	6.9	6.0	6.5	14.0	5.7	4.5	4.0
JB-1a	pref.	7.1	26.15	5.099	1.484	4.7	0.699	4.07	0.805	2.232	0.3197	2.1	0.3147	3.47	1.738	0.072	6.44	8.97	1.615
JB-1b	ET14	6.84	24.90	4.73	1.34	4.25	0.67	3.70	0.72	2.04	0.28	1.85	0.27	3.09	1.34	0.058	5.31	8.83	1.47
	ET27	7.21	26.08	5.04	1.43	4.64	0.73	3.97	0.81	2.18	0.32	2.11	0.31	3.51	1.55	0.080	5.10	9.52	1.53
	IZ11	6.75	24.89	4.67	1.33	4.25	0.67	3.75	0.72	2.08	0.28	1.81	0.26	3.15	1.35	0.054	5.33	8.67	1.50
	EG52	6.74	24.93	4.63	1.33	4.21	0.67	3.69	0.73	2.05	0.29	1.83	0.28	3.07	1.36	0.071	5.51	8.76	1.53
	average	6.88	25.20	4.77	1.36	4.34	0.69	3.78	0.75	2.09	0.29	1.90	0.28	3.21	1.40	0.07	5.31	8.94	1.51
JB-1b	RSD	0.22	0.59	0.18	0.05	0.20	0.03	0.13	0.04	0.06	0.02	0.14	0.02	0.21	0.10	0.01	0.17	0.39	0.03
	RSD (%)	3.2	2.3	3.9	3.5	4.6	4.5	3.5	5.8	3.1	7.0	7.5	8.5	6.5	7.2	18.1	3.1	4.3	1.8
	recom.																6.8		
JB-2	KH02B	1.07	5.92	2.17	0.76	2.98	0.55	3.66	0.80	2.40	0.35	2.31	0.36	1.46	0.029	0.034	4.93	0.26	0.13
	EG24	1.09	6.03	2.17	0.76	2.99	0.55	3.71	0.81	2.46	0.36	2.36	0.36	1.46	0.035	0.032	4.20	0.21	0.12
	EG36	1.07	5.95	2.13	0.77	2.98	0.55	3.68	0.81	2.43	0.35	2.32	0.35	1.40	0.031	0.031	3.66	0.19	0.11
	OJ35	0.98	5.42	1.96	0.70	2.82	0.50	3.39	0.74	2.23	0.32	2.14	0.33	1.40	0.034	0.022	4.05	0.21	0.12
	KF37	1.04	5.77	2.09	0.73	2.86	0.53	3.54	0.77	2.32	0.33	2.24	0.34	1.46	0.031	0.029	4.19	0.23	0.14
	PP12	1.04	5.72	2.05	0.72	2.82	0.52	3.49	0.76	2.29	0.33	2.20	0.33	1.32	0.030	0.029	4.37	0.23	0.14
	OJ48	1.06	5.90	2.13	0.75	3.00	0.54	3.67	0.81	2.42	0.35	2.34	0.36	1.44	0.035	0.033	4.60	0.22	0.12
	OJ59	1.07	5.96	2.14	0.75	3.02	0.55	3.69	0.81	2.44	0.35	2.36	0.36	1.45</					

Table 3 (continued). Trace element concentrations ($\mu\text{g/g}$; ppm) in reference rocks.

Sample	Run No.	Pr	Nd	Sm	Eu	Gd	Tb	Dy	Ho	Er	Tm	Yb	Lu	Hf	Ta	Tl	Pb	Th	U	
JA-1	KH04B	2.09	10.69	3.36	1.07	4.14	0.72	4.62	0.99	3.00	0.43	2.81	0.45	2.54	0.086	0.082	6.12	0.71	0.30	
	PP28	1.98	10.22	3.23	1.05	4.12	0.71	4.62	0.98	2.94	0.42	2.83	0.43	2.47	0.090	0.079	5.60	0.75	0.35	
	PA15	2.11	10.80	3.38	1.06	4.27	0.73	4.62	0.99	2.93	0.43	2.85	0.43	2.46	0.086	0.080	5.16	0.70	0.33	
	A15	2.04	10.53	3.33	1.06	4.25	0.74	4.67	1.02	2.96	0.44	2.98	0.45	2.61	0.202	0.093	5.21	0.76	0.35	
	ET28	2.11	10.76	3.43	1.09	4.45	0.76	4.81	1.05	3.05	0.46	3.11	0.48	2.71	0.098	0.097	5.47	0.82	0.37	
	A47	2.04	10.49	3.30	1.06	4.16	0.72	4.65	0.99	2.96	0.43	2.81	0.43	2.52	0.091	0.081	5.20	0.71	0.34	
JA-1	average	2.06	10.58	3.34	1.07	4.23	0.73	4.67	1.01	2.97	0.44	2.90	0.45	2.55	0.11	0.085	5.46	0.74	0.34	
	RSD	0.05	0.22	0.07	0.01	0.12	0.02	0.08	0.03	0.04	0.02	0.12	0.02	0.10	0.05	0.008	0.37	0.04	0.02	
JA-1	pref.	2.082	10.69	3.396	1.112	4.15	0.727	4.75	1.032	2.959	0.445	2.949	0.454	2.51	0.0979	0.106	5.86	0.761	0.34	
JA-2	MG08	3.54	13.67	2.90	0.82	2.79	0.46	2.77	0.56	1.64	0.23	1.54	0.23	2.80	0.57	0.30	14.24	3.41	1.57	
	ET29	3.72	14.17	3.08	0.88	3.04	0.51	2.94	0.62	1.76	0.27	1.77	0.27	3.21	0.67	0.38	13.37	3.76	1.66	
	ET30	3.78	14.36	3.13	0.89	3.10	0.51	2.94	0.63	1.75	0.27	1.77	0.27	3.25	0.68	0.38	13.14	3.71	1.63	
	IZ13	3.56	13.86	2.95	0.83	2.88	0.47	2.81	0.57	1.69	0.23	1.56	0.23	2.92	0.60	0.33	18.38	4.58	2.11	
JA-2	average	3.65	14.01	3.02	0.85	2.96	0.49	2.87	0.59	1.71	0.25	1.66	0.25	3.04	0.63	0.35	14.78	3.86	1.74	
	RSD	0.12	0.31	0.11	0.04	0.14	0.02	0.09	0.04	0.06	0.02	0.13	0.02	0.22	0.06	0.04	2.45	0.50	0.25	
JA-2	pref.	3.691	14.04	3.032	0.893	3.013	0.4786	2.851	0.591	1.676	0.2546	1.645	0.2549	2.838	0.652	0.33	18.88	4.8	2.182	
JA-3	KH05B	2.86	12.63	3.22	0.80	3.23	0.54	3.34	0.69	2.08	0.30	1.98	0.32	3.41	0.23	0.20	9.97	3.02	0.88	
	ET15	2.84	12.32	3.10	0.76	3.25	0.54	3.25	0.68	2.01	0.29	1.98	0.30	3.26	0.23	0.19	7.18	3.07	0.92	
	ET31	2.82	12.28	3.11	0.78	3.38	0.56	3.34	0.73	2.09	0.32	2.14	0.33	3.58	0.25	0.23	6.50	3.08	0.89	
	AM15	2.72	11.95	3.03	0.73	3.16	0.53	3.22	0.68	2.01	0.30	1.96	0.29	3.41	0.25	0.20	6.19	2.73	0.83	
	A31	2.69	11.63	2.89	0.69	2.95	0.50	3.05	0.63	1.87	0.27	1.82	0.27	2.89	0.20	0.18	6.30	2.69	0.80	
	average	2.78	12.16	3.07	0.75	3.20	0.53	3.24	0.68	2.01	0.30	1.97	0.30	3.31	0.23	0.20	7.23	2.92	0.86	
	RSD	0.08	0.38	0.12	0.04	0.16	0.02	0.12	0.03	0.09	0.02	0.11	0.02	0.26	0.02	0.02	1.58	0.19	0.05	
JA-3	pref.	2.40	12.30	3.05	0.82	2.96	0.52	3.01	0.51	1.57	0.28	2.16	0.32	3.42	0.27	0.230	7.7	3.25	1.18	
JR-2	KH06B	4.91	19.40	5.42	0.11	5.80	1.06	6.72	1.44	4.70	0.75	5.23	0.85	5.35	2.00	1.86	22.2	28.9	9.65	
	IZ14	4.57	18.20	5.02	0.09	5.26	0.98	6.32	1.35	4.33	0.68	4.93	0.76	4.89	1.85	1.69	19.4	28.1	9.82	
	P31	4.60	18.39	5.09	0.09	5.38	1.00	6.49	1.39	4.43	0.71	5.09	0.78	5.01	1.91	1.77	20.3	29.9	10.44	
	EG54	4.54	18.13	4.98	0.09	5.30	0.98	6.29	1.37	4.31	0.69	4.95	0.77	4.88	1.87	1.70	19.6	28.5	9.96	
	average	4.66	18.53	5.13	0.09	5.44	1.01	6.46	1.39	4.45	0.71	5.05	0.79	5.03	1.91	1.75	20.34	28.83	9.97	
	RSD	0.17	0.59	0.20	0.01	0.25	0.04	0.20	0.04	0.18	0.03	0.14	0.04	0.22	0.07	0.08	1.27	0.77	0.34	
JR-2	recom.	4.75	20.40	5.63	0.14	5.83	1.10	6.63	1.39	4.36	0.74	5.33	0.88	5.14	2.29	1.850	21.5	31.4	10.90	
BIR-1a	KH07B	0.37	2.37	1.11	0.52	1.83	0.35	2.51	0.56	1.66	0.24	1.59	0.25	0.62	0.027	0.0050	2.97	0.070	0.014	
	KH33	0.34	2.14	0.99	0.46	1.64	0.32	2.27	0.51	1.51	0.22	1.44	0.22	0.54	0.026	0.0026	2.54	0.042	0.009	
	KH59	0.37	2.38	1.12	0.52	1.84	0.37	2.59	0.58	1.74	0.25	1.67	0.25	0.63	0.033		2.68	0.024	0.006	
	KH71	0.35	2.29	1.06	0.50	1.76	0.35	2.52	0.55	1.67	0.24	1.57	0.24	0.59	0.037		3.04	0.029	0.010	
	KH83	0.33	2.14	0.99	0.46	1.67	0.33	2.29	0.51	1.53	0.22	1.47	0.22	0.53	0.032	0.0011	2.59	0.056	0.027	
	OR11	0.33	2.16	1.00	0.48	1.70	0.34	2.36	0.53	1.60	0.23	1.51	0.23	0.53		0.0001	2.62	0.060	0.004	
	OR23	0.34	2.18	1.02	0.48	1.71	0.34	2.43	0.54	1.63	0.24	1.57	0.24	0.55	0.020	0.0021	2.75	0.054	0.011	
	AF12	0.34	2.16	1.01	0.47	1.70	0.33	2.37	0.53	1.59	0.23	1.52	0.23	0.55	0.025		2.75	0.031	0.002	
	AF24	0.36	2.27	1.07	0.50	1.78	0.35	2.49	0.56	1.68	0.24	1.59	0.24	0.59	0.035	0.0006	2.70	0.029	0.005	
	EG12	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	2.83	0.030	0.010	
	KI15	0.36	2.29	1.06	0.51	1.85	0.35	2.49	0.56	1.67	0.24	1.62	0.25	0.60	0.032	0.0021	3.23	0.029	0.008	
	BIR-1a	average	0.35	2.25	1.05	0.49	1.75	0.35	2.44	0.54	1.64	0.24	1.56	0.24	0.58	0.03	0.0019	2.79	0.041	0.010
		RSD	0.01	0.09	0.05	0.02	0.07	0.01	0.11	0.02	0.07	0.01	0.07	0.01	0.04	0.01	0.0015	0.21	0.016	0.007
		RSD (%)	4.1	4.2	4.7	4.3	4.2	4.2	4.3	4.4	4.7	4.5	4.5	6.2	20.1	76.5	7.5	38.5	69.4	
	BIR-1	pref.	0.3723	2.397	1.113	0.5201	1.809	0.3623	2.544	0.5718	1.68	0.2558	1.631	0.2484	0.5822	0.0414	0.0021	3.037	0.0328	0.01051
BHVO-2	SA7	5.02	23.11	5.69	1.90	5.69	0.87	4.88	0.88	2.32	0.29	1.79	0.24	3.92	0.87		1.38	1.14	0.38	
	SA33	4.82	22.37	5.46	1.87	5.67	0.88	4.79	0.91	2.34	0.31	1.82	0.26	4.18	1.02	0.025	1.21	1.07	0.38	
	IZ15	4.96	22.87	5.66	1.89	5.78	0.89	4.96	0.91	2.39	0.29	1.83	0.25	4.25	0.97	0.015	1.47	1.15	0.39	
	SA49	5.06	23.46	5.85	1.94	5.89	0.91	5.05	0.92	2.45	0.31	1.92	0.26	4.32	1.01	0.010	1.33	1.10	0.36	
	SA65	5.09	23.65	5.86	1.96	5.90	0.90	5.08	0.93	2.47	0.30	1.90	0.26	4.33	1.06	0.014	1.51	1.16	0.38	
	average	4.99	23.09	5.71	1.91	5.79	0.89	4.95	0.91	2.39	0.30	1.85	0.25	4.20	0.99	0.016	1.38	1.12	0.38	
	RSD	0.11	0.50	0.16	0.04	0.11	0.02	0.12	0.02	0.07	0.01	0.05	0.01	0.17	0.07	0.006	0.12	0.04	0.01	
	RSD (%)	2.1	2.2	2.9	1.9	1.9	1.7	2.4	2.2	2.8	3.6	2.9	4.1	4.0	7.4	39.7	8.5	3.4	3.2	
BHVO-2	pref.	5.339	24.27	6.023	2.043	6.207	0.9392	5.28	0.9887	2.511	0.3349	1.994	0.2754	4.47	1.154	0.0224	1.653	1.224	0.412	
JSd-1	E16	3.95	15.45	3.26	0.83	2.86	0.42	2.33	0.44	1.24	0.16	1.08	0.15	0.74	0.64	0.29	8.97	2.87	0.58	
	SA17	3.87	15.58	3.30	0.87	2.94	0.46	2.48	0.48	1.32	0.18	1.17	0.18	0.91	0.70	0.34	9.19	2.98	0.61	
	PI15	4.03	16.26	3.41	0.91	3.14	0.48	2.60	0.50	1.38	0.19	1.22	0.19	0.92	0.75	0.36	9.60	3.11	0.66	
	KM15	3.89	15.49	3.26	0.86	2.90	0.44	2.41	0.47	1.29	0.18	1.14	0.16	0.82	0.70	0.31	9.60	2.99	0.63	
	average	3.93	15.69	3.31	0.87	2.96</														

Table 4. Selected major element concentrations (wt%) in reference rocks.

Sample	Run No.	TiO ₂	Al ₂ O ₃	Fe ₂ O ₃ *	MnO	MgO	CaO	Na ₂ O	K ₂ O	P ₂ O ₅
JB-1	KH34m	1.383	14.11	8.55	0.149	7.70	9.10	2.74	1.40	0.267
	KH70m	1.384	14.78	8.72	0.153	7.61	9.10	2.73	1.39	0.269
	OR22m	1.384	14.51	8.73	0.148	7.68	9.11	2.74	1.34	0.269
	ES16	1.378	15.28	9.10	0.153	8.31	8.99	2.77	1.41	0.287
	ET48	1.381	15.29	8.90	0.162	8.16	8.85	2.79	1.41	0.279
	KM14-2	1.359	14.77	8.80	0.153	7.66	9.03	2.66	1.39	0.305
JB-1	average	1.378	14.79	8.80	0.153	7.85	9.03	2.74	1.39	0.279
	RSD**	0.009	0.42	0.17	0.004	0.27	0.09	0.04	0.02	0.013
	RSD (%)#	0.7	2.8	2.0	2.8	3.5	1.0	1.5	1.8	4.8
JB-1	pref.#	1.299	14.53	8.99	0.149	7.807	9.33	2.749	1.42	0.263
JB-1a	KF36m	1.368	14.32	8.98	0.146	7.68	9.27	2.71	1.38	0.264
	PP11m	1.370	14.18	8.89	0.149	7.64	9.24	2.70	1.38	0.266
	OJ47m	1.515	15.51	9.85	0.162	8.46	10.22	3.00	1.52	0.289
	OJ90m	1.365	14.15	8.96	0.147	7.72	9.21	2.72	1.39	0.263
	PP27m	1.385	14.36	8.95	0.147	7.82	9.30	2.77	1.40	0.271
	KI14m	1.395	14.37	8.98	0.149	7.89	9.36	2.76	1.40	0.274
	PA14m	1.386	14.14	8.98	0.148	7.73	9.24	2.73	1.40	0.262
	A14m	1.336	13.67	8.71	0.145	7.37	9.10	2.64	1.35	0.255
	A30m	1.331	13.73	8.79	0.145	7.35	9.13	2.66	1.35	0.253
	SA6m	1.338	13.76	8.81	0.145	7.51	9.03	2.65	1.35	0.258
	SA16m	1.325	13.66	8.73	0.144	7.52	9.00	2.65	1.35	0.252
	AM14m	1.354	13.87	8.89	0.146	7.62	9.17	2.66	1.37	0.259
	PI14	1.358	14.72	8.90	0.148	7.81	9.03	2.70	1.37	0.299
	SA64	1.357	14.78	8.96	0.149	7.86	9.09	2.72	1.39	0.309
	KM14	1.354	14.78	8.90	0.149	7.82	9.13	2.65	1.39	0.306
	EG51	1.393	15.32	9.17	0.152	8.23	9.45	2.77	1.42	0.266
JB-1a	average	1.371	14.33	8.97	0.148	7.75	9.25	2.72	1.39	0.272
	RSD	0.044	0.57	0.26	0.004	0.29	0.29	0.09	0.04	0.019
	RSD (%)	3.2	4.0	2.9	2.9	3.7	3.1	3.3	3.0	6.9
JB-1a	pref.	1.29	14.51	8.996	0.1419	7.81	9.314	2.738	1.407	0.261
JB-1b	IZ11	1.352	14.96	9.12	0.148	8.78	9.77	2.71	1.34	0.261
	ET14m	1.287	13.51	8.64	0.141	8.07	9.26	2.53	1.24	0.255
	ET27m	1.279	13.45	8.72	0.141	7.97	9.29	2.51	1.24	0.250
	IZ11m	1.322	14.51	9.00	0.146	8.22	9.54	2.57	1.29	0.249
	EG52	1.335	15.04	9.12	0.147	8.66	9.56	2.66	1.32	0.264
JB-1b	average	1.315	14.29	8.92	0.145	8.34	9.48	2.57	1.29	0.26
	RSD	0.031	0.77	0.23	0.003	0.36	0.21	0.07	0.05	0.01
	RSD (%)	2.4	5.4	2.5	2.2	4.3	2.2	2.6	3.6	2.6
JB-1b	recom.##	1.26	14.38	9.02	0.147	8.14	9.60	2.63	1.32	0.256
JB-2	OJ35m	1.106	13.37	12.82	0.197	3.98	8.73	1.78	0.366	0.088
	PP12m	1.203	14.49	13.96	0.215	4.37	9.49	1.94	0.398	0.095
	OJ48m	1.220	14.64	14.00	0.217	4.41	9.51	1.95	0.401	0.095
	OJ59m	1.235	14.86	14.14	0.217	4.51	9.64	1.99	0.408	0.095
JB-2	average	1.191	14.341	13.729	0.212	4.32	9.34	1.92	0.393	0.093
	RSD	0.058	0.663	0.613	0.010	0.23	0.42	0.09	0.019	0.004
	RSD (%)	4.9	4.6	4.5	4.6	5.3	4.4	4.8	4.7	3.8
JB-2	pref.	1.167	14.62	14.28	0.213	4.43	9.852	2.054	0.4224	0.097
JB-3	OJ76m	1.521	17.63	11.81	0.181	5.10	9.69	2.69	0.77	0.304
	OJ91m	1.504	17.40	11.70	0.180	5.01	9.64	2.67	0.76	0.294
	EG53	1.536	18.40	12.06	0.186	5.34	9.94	2.81	0.79	0.312
JB-3	average	1.520	17.81	11.86	0.182	5.15	9.76	2.72	0.77	0.30
	RSD	0.016	0.52	0.18	0.003	0.17	0.16	0.08	0.02	0.01
	RSD (%)	1.0	2.9	1.5	1.8	3.3	1.7	2.8	2.3	2.9
JB-3	recom.	1.44	17.20	11.82	0.177	5.19	9.79	2.73	0.78	0.294

Sample	Run No.	TiO ₂	Al ₂ O ₃	Fe ₂ O ₃ *	MnO	MgO	CaO	Na ₂ O	K ₂ O	P ₂ O ₅
JA-1	PP28m	0.89	14.65	6.96	0.156	1.41	5.61	3.75	0.755	0.158
	PA15m	0.89	14.59	6.90	0.156	1.39	5.63	3.74	0.754	0.155
	A15m	0.87	14.44	6.82	0.153	1.37	5.49	3.68	0.738	0.154
	ET28m	0.88	14.54	6.85	0.154	1.37	5.56	3.73	0.745	0.164
	A47	0.88	15.78	6.95	0.156	1.58	5.55	3.86	0.758	0.177
JA-1	average	0.882	14.80	6.90	0.155	1.42	5.57	3.75	0.750	0.161
	RSD	0.009	0.55	0.06	0.002	0.09	0.06	0.07	0.009	0.009
JA-1	pref.	0.85	15.19	7.05	0.1543	1.54	5.72	3.91	0.779	0.1595
JA-2	IZ13	0.74	16.90	6.41	0.110	7.60	6.47	3.050	1.94	0.163
	ET29m	0.701	14.92	6.140	0.107	6.53	6.09	3.069	1.76	0.159
	ET30m	0.704	15.00	6.095	0.105	6.45	6.09	3.069	1.76	0.154
	IZ13m	0.720	16.43	6.312	0.108	7.15	6.30	3.148	1.84	0.151
JA-2	average	0.716	15.81	6.24	0.11	6.93	6.24	3.09	1.83	0.157
	RSD	0.017	1.00	0.15	0.00	0.54	0.18	0.05	0.09	0.005
JA-2	pref.	0.670	15.51	6.289	0.109	7.84	6.26	3.072	1.779	0.152
JA-3	ET15m	0.71	15.43	6.51	0.105	3.43	6.15	3.06	1.36	0.112
	ET31m	0.71	15.23	6.48	0.105	3.39	6.22	3.10	1.37	0.109
	A31m	0.74	16.21	6.62	0.107	3.65	6.37	3.15	1.40	0.115
JA-3	average	0.720	15.62	6.53	0.106	3.49	6.25	3.10	1.38	0.112
	RSD	0.015	0.52	0.07	0.001	0.14	0.11	0.04	0.03	0.003
JA-3	recom.	0.70	15.56	6.60	0.104	3.72	6.24	3.19	1.41	0.116
JR-2	IZ14	0.063	13.94	0.78	0.117	0.042	0.56	4.909	0.004	
	IZ14m	0.056	13.19	0.76	0.113	0.038	0.53	4.02	4.648	0.003
	EG54	0.064	13.29	0.76	0.112	0.043	0.56	4.05	4.481	0.007
JR-2	average	0.061	13.47	0.77	0.114	0.041	0.55	4.03	4.679	0.005
	RSD	0.005	0.41	0.02	0.003	0.002	0.02	0.216	0.002	
JR-2	recom.	0.07	12.72	0.77	0.112	0.04	0.50	3.99	4.45	0.012
BIR-1a	KH59m	1.077	15.80	11.78	0.181	9.925	13.64	1.894	0.023	0.026
	BIR-1a	1.024	15.07	11.36	0.173	9.528	13.05	1.788	0.021	0.025
	OR23m	1.071	15.69	11.97	0.184	10.017	13.75	1.873	0.022	0.026
	BIR-1a	1.038	15.05	11.35	0.174	9.477	13.10	1.796	0.022	0.025
	BIR-1a	1.011	14.84	11.34	0.174	9.338	12.91	1.767	0.021	0.025
BIR-1a	average	1.044	15.29	11.56	0.177	9.66	13.29	1.824	0.022	0.025
	RSD	0.029	0.43	0.30	0.005	0.30	0.38	0.056	0.001	0.001
	RSD (%)	2.8	2.8	2.6	2.8	3.1	2.8	3.1	4.9	2.8
BIR-1	pref.	0.959	15.51	11.40	0.173	9.689	13.29	1.832	0.029	0.03
BHVO-2	IZ15	3.042	14.74	13.11	0.181	7.80	11.98	2.22	0.563	0.282
	BHVO-2	2.888	13.97	12.46	0.174	7.38	11.28	2.22	0.528	0.333
	SA7	2.866	14.32	12.61	0.175	7.44	11.46	2.21	0.530	0.321
	SA49	2.986	14.38	12.74	0.176	7.53	11.49	2.25	0.532	0.332
	SA65	2.782	13.47	12.23	0.166	7.09	10.76	2.12	0.502	0.298
	SA33									
BHVO-2	average	2.913	14.18	12.63	0.174	7.45	11.39	2.20	0.531	0.313
	RSD	0.102	0.48	0.33	0.005	0.26	0.44	0.05	0.022	0.022
	RSD (%)	3.5	3.4	2.6	3.1	3.4	3.9	2.5	4.1	7.1
BHVO-2	pref.	2.731	13.44	12.39	0.169	7.527	11.4	2.219	0.513	0.269
JSd-1	PI15	0.681	15.22	5.06	0.092	1.79	3.03	2.73	2.162	0.135
	KM15	0.663	14.91	4.94	0.089	1.76	2.94	2.69	2.105	0.136
JSd-1	average	0.672	15.06	5.00	0.090	1.77	2.98	2.71	2.133	0.136
JSd-1	recom.	0.643	14.65	5.059	0.092	1.813	3.034	2.727	2.183	0.122

* Total iron as Fe₂O₃.

** One sigma standard deviation of the replicated measurements.

RSD (%) of only basaltic samples are shown because other samples have small number of a replicated analyses.

GeoRem preferred values (<http://georem.mpch-mainz.gwdg.de/>) are shown in green.## Recommended, certified, or information values (<https://gbank.gsj.jp/geostandards/welcome.html>) are also shown in green.

tions, was present in the sedimentary material. If zircon is present in a sample rock, we cannot use Zr and Hf data obtained from acid digested solutions for geological discussion. Finally, we note that REE concentrations analyzed in this study are slightly (1–7%) lower than the reference values for basaltic rocks, but the slight differences would not affect our geological discussion.

Summary

We established routine sample digestion procedures suitable for ICP-MS measurements of volcanic rocks and sediments, and we quantitatively analyzed 37 trace and 9 selected major element concentrations in 12 geological standard rocks. The reproducibility of replicated analyses (i.e., measurement of several different solutions of the same reference material) in the 12 geological standard rocks is better than 5% RSD for almost of all elements. Our measured concentrations of most major and trace elements in the 12 geological standard rocks agree with the reference values.

Acknowledgements

This research was financially supported by Japan Society for the Promotion of Science (JSPS) KAKENHI Grant Numbers 18H05447, 18H03746, and the project of the National Museum of Nature and Science, Chemical Stratigraphy and Dating as a Clue for Understanding the History of the Earth and Life. We thank Ryo Amma for providing the JSPS budget (18H05447), Ginta Motohashi, Taishi

Suzuki, Kento Motomura, Naoto Kimura and Kohei Ishikawa for analytical support, and Ryoko Senda for a peer review.

References

- Chang, Q., Shibata, T., Shinotsuka, K., Yoshikawa, M. and Tatsumi, Y. (2003) Precise determination of trace elements in geological standard rocks using inductively coupled plasma mass spectrometry. *Frontier Research on Earth Evolution (IFREE Report for 2001–2002)*, **1**: 357–362.
- May, T. W. and Wiedmeyer, R. H. (1998) A table of polyatomic interferences in ICP-MS. *Atomic Spectroscopy*, **19**: 150–155.
- Nakamura, K. and Chang, Q. (2007) Precise determination of ultra-low (sub-ng g⁻¹) level rare earth elements in ultramafic rocks by Quafupole ICP-MS. *Geostandards and Geoanalytical Research*, **31**: 185–197.
- Pretorius, W., Weis, D., Williams, G., Hanano, D., Kieffer, B. and Scoates, J. (2003) Complete trace element characterization of granitoid (USGS G-2, GSP-2) reference materials by high resolution inductively coupled plasma-mass spectrometry. *Geostandards and Geoanalytical Research*, **30**: 39–54.
- Sano, T., Shirao, M., Tani, K., Tsutsumi, Y., Kiyokawa, S. and Fujii, T. (2016) Progressive enrichment of arc magmas caused by the subduction of seamounts under Nishinoshima volcano, Izu–Bonin Arc, Japan. *Journal of Volcanology and Geothermal Research*, **319**: 52–65.
- Sano, T., Tani, K. and Murch, A. P. (2020). Major and trace element analyses of igneous rocks and sediments by X-ray fluorescence spectrometry using glass bead and pressed powder pellet. *Bulletin of the National Museum of Nature and Science, Series C*, **46**: 1–14.
- Senda, R., Kimura, J.-I. and Chang, Q. (2014) Evaluation of a rapid, effective sample digestion method for trace element analysis of granitoid samples containing acid-resistant materials: Alkali fusion after acid digestion. *Geochemical Journal*, **48**: 99–103.

Appendix: A procedure for acid digestion of rock powder

1. Left 0.5M HNO₃ out of PFA vial and clean the vial and cap with Milli-Q water (twice).
2. Dry on a heater (P-4: 80°C, ~1 h)
3. Write run number on the PFA vial and cap
4. Weigh (0.0500–0.0550 g) sample powder and put in the PFA vial
5. Add 666 ml HClO₄ [blue pipette scale = 0.666]
6. Add 2,000 ml HF in draft chamber [blue pipette scale = 1.000 × 2] and close cap tightly
7. Ultrasonic cleaning (20 min)
8. Heat (P-1: 150°C 15h, 160°C 20h) and wait 1 day 8 h
9. Tap the PHA vial to dislodge droplets
10. Dryness: Open cap and dry on a heater (P2: 80°C 20h, 110°C 10h, 150°C 15h, 160°C 20–30h) and wait 2 days 17 h~3 days 3 h
11. Close cap (during T = 160°C) and switch off the heater and wait until T decreases (<40°C)
12. Add 1,000 ml HClO₄ (blue pipette scale = 1.000) and close cap tightly
13. Heat (P-3: 160°C 12 h)
14. Add 2,000 ml 6M HNO₃ (blue pipette scale = 1.000 × 2) and close cap tightly
15. Heat (P-4: 80°C ~1 h)
16. Dryness: Open cap and dry on a heater (P2: 80°C 20h, 110°C 10h, 150°C 15h, 160°C 20–30h) and wait 2 days 17 h~3 days 3 h
(Left Milli-Q water out of 15 ml and 10 ml bottles and dry in clean booth)
17. Close cap (during T = 160°C) and switch off the heater and wait until T decreases (<40°C)
18. Write run number on 15 ml bottle and weigh (~7.3 g)
19. Add ~7.5 ml of 2% HNO₃ + 0.1% HF in PFA jar and close cap, and shake well, and pour the solution into the 15 ml bottle [two times]
20. Weigh the solution in the 15 ml bottle (~20–23 g) [dilution factor: ~300]
21. Write run number on 10 ml bottle and weigh (~2.8 g)
22. Add 100–500 μl of the solution [by using blue pipette] in the 10 ml bottle and weigh (e.g., JB-1a = 100 μl, BIR-1a = 500 μl, BHVO = 100 μl, JB-2 = 200 μl, JB-3 = 100 μl, blank = 1,000 μl)
23. Add 200 μl Internal standard (In & Bi~100 ppb) [by using yellow pipette] and weigh
24. Add 2% HNO₃ + 0.1% HF up to 10 ml and weigh